

EU Type Examination Certificate

This is to certify that:

CleanSpace Technology Pty Ltd
U5, 39 Herbert Street
St. Leonards
New South Wales
2065
Australia

Holds Certificate Number:

CE 704321

In respect of:

**Respiratory Protective Devices
To EN 12942:1998 + A2:2008**

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Denelise L'Ecluse, Managing Director Assurance - Continental Europe

First Issued: 2019-04-02
Latest Issue: 2024-08-26

Effective Date: 2024-08-26
Expiry Date: 2029-04-02



EU Type Examination Certificate

No. CE 704321

Product Specification

Product Name:	CleanSpace
Product Type:	Range of power assisted filtering devices incorporating full face masks or half masks
Technical Specification:	EN 12942:1998+A2:2008 Respiratory protective devices — Power assisted filtering devices incorporating full face masks, half masks or quarter masks
Product Description:	The CleanSpace range of respiratory products are head mounted full and half masks to EN 12942:1998+A2:2008 as detailed below;

Certified Systems:

CleanSpace2™ System

Part No.	Component	Product description
PAF-0034	CleanSpace2 Blower Unit	Classification - TM3
PAF-0033	H-Series Mask - Small	MMDF - 120 l/min
PAF-1010	H-Series Mask - Medium	Duration - 4 hours
PAF-0027	H-Series Mask - Large	Filters; Particulate
PAF-0019	Q-Series Mask - Medium	A1 P SL R
PAF-0020	Q-Series Mask - Large	A2 P SL R
PAF-0035	Filter TM3	ABE1 P SL R
PAF-0038	Filter Adaptor	ABEK1 P SL R
PAF-0037	Filter TM3 High Capacity	
PAF-0050	Filter TM3 A1 P SL R	
PAF-0077	Filter TM3 A2 P SL R	Classification - TM2
PAF-0051	Filter TM3 ABE1P SL R	MMDF - 120 l/min
PAF-0052	Filter TM3 ABEK1 P SL R	Duration - 4 hours
PAF-0087	Filter TM2AXP	Filters AX P
CS3038	Steri-plus exhalation filter (Case)	
CS3039	Steri-plus exhalation filter (filter)	

First Issued: 2019-04-02

Latest Issue: 2024-08-26

Effective Date: 2024-08-26

Expiry Date: 2029-04-02

Page: 2 of 6

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

EU Type Examination Certificate

No. CE 704321

Product Specification (continued)

Certified Systems:	Part No.	Component	Product description
CleanSpace ULTRA™ System	PAF-0071	CleanSpace ULTRA Blower Unit	Classification - TM3
	PAF-0033	H-Series Mask - Small	MMDF - 120 l/min
	PAF-1010	H-Series Mask - Medium	Duration - 4 hours
	PAF-0027	H-Series Mask - Large	Filters; Particulate
	PAF-1014	Full Face Mask - Medium/Large	A1 P SL R
	PAF-1106	Full Face Mask – Small	A2 P SL R
	PAF-0035	Filter TM3	ABE1 P SL R
	PAF-0038	Filter Adaptor	ABEK1 P SL R
	PAF-0037	Filter TM3 High Capacity	Classification - TM2
	PAF-0050	Filter TM3 A1 P SL R	MMDF - 120 l/min
	PAF-0077	Filter TM3 A2 P SL R	Duration - 4 hours
	PAF-0051	Filter TM3 ABE1 P SL R	Filters AX P
	PAF-0052	Filter TM3 ABEK1 P SL R	
	PAF-0087	Filter TM2AXP	
	CS3038	Steri-plus exhalation filter (Case)	
	CS3039	Steri-plus exhalation filter (filter)	

First Issued: 2019-04-02
Latest Issue: 2024-08-26

Effective Date: 2024-08-26
Expiry Date: 2029-04-02

Page: 3 of 6

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 704321

Product Specification (continued)

Certified Systems:	Part No.	Component	Product description
CleanSpace ULTRA™ & EX Systems	PAF-0061	CleanSpace EX Blower Unit	Classification - TM3
	PAF-0033	H-Series Mask - Small	MMDF - 120 l/min
	PAF-1010	H-Series Mask - Medium	Duration - 4 hours
	PAF-0027	H-Series Mask - Large	Filters; Particulate
	PAF-0064	H-Series Mask EX - Small**	A1 P SL R
	PAF-0062	H-Series Mask EX - Medium**	A2 P SL R
	PAF-0063	H-Series Mask EX - Large**	ABE1 P SL R
	PAF-1014	Full Face Mask - Medium/Large	ABEK1 P SL R
	PAF-1106	Full Face Mask – Small	
	PAF-0035	Filter TM3	Classification - TM2
	PAF-0078	Filter Adaptor - EX	MMDF - 120 l/min
	PAF-0037	Filter TM3 High Capacity	Duration - 4 hours
	PAF-0050	Filter TM3 A1	Filters AX P
	PAF-0077	Filter TM3 A2	
	PAF-0051	Filter TM3 ABE1	
	PAF-0052	Filter TM3 ABEK1	
	PAF-0087	Filter TM2AXP	
	CS3038	Steri-plus exhalation filter (Case)	
	CS3039	Steri-plus exhalation filter (filter)	

First Issued: 2019-04-02

Latest Issue: 2024-08-26

Effective Date: 2024-08-26

Expiry Date: 2029-04-02

Page: 4 of 6

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

EU Type Examination Certificate

No. CE 704321

Product Specification (continued)

Certified Systems:	Part No.	Component	Product description
CleanSpace HALO™ System	CS3000	CleanSpace Halo Blower Unit	Classification - TM3
	CS3003	Half Mask - Small	MMDF - 120 l/min
	CS3004	Half Mask - Medium	Duration - 6 hours
	CS3005	Half Mask - Large	Filters; Particulate
	CS3002	Filter TM3 P R SL	
	CS3006	Full facemask with harness – HALO Small	
	CS3007	Full facemask with harness – HALO Medium/Large	
	CS3024	Headcovers – HALO BIO	
	CS3025	Filter HE / TM3 P R SL – HALO BIO	
	CS3026	BIO Exhalation Valve Cover	
	CS3027	Exhalation Filter – HALO BIO	
	CS3038	Steri-plus exhalation filter (Case)	
	CS3039	Steri-plus exhalation filter (filter)	

First Issued: 2019-04-02
Latest Issue: 2024-08-26

Effective Date: 2024-08-26
Expiry Date: 2029-04-02

Page: 5 of 6

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 704321

Certificate Administration Details

Technical File Reference: Cleanspace Technical File

Certificate Amendment Record:

Issue date	Comments	BSI review No.
August 2024	5yr renewal; update Module D reference; clarify PAF-0087 TM2 classification.	2797:24:30103504
September 2022	Addition of Steri-plus exhalation filters CS3038 & CS3039	2797:22:3563875
April 2022	Addition of CS3006, CS3007, CS3024, CS30025, CS3026, CS3027 & PAF-0087	2797:22:3090237
April 2019	First issue to PPE Regulation 2016/425. Products previously certified to PPE Directive 89/686 on BSI certificate CE 682864	0086:19:9688463

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The Conformity to Type Based on Quality Assurance of the Production Process under Module D for the product are referenced in CE Certificate number 0082/3959/D/07/24/0066 issued by Notified Body Apave (NB 0082).

First Issued: 2019-04-02

Latest Issue: 2024-08-26

Effective Date: 2024-08-26

Expiry Date: 2029-04-02

Page: 6 of 6

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.